

Desinfeksjonsmidler som anbefales til ultralydprober og som produsentene har bekreftet kan benyttes uten å skade deres prober. November 2024			Desinfeksjonsmidler til teknisk bruk Bruksområder, effekt og virketider for desinfeksjonsmidler Infeksjonskontroll.no og desinfeksjonskluter Infeksjonskontroll.no		
Produsent av ultralydprober	Varemerker	Leverandør	Pereddiksyre	Kloridioksid	Etanol brukes bare under rene forhold
Siemens Healthineers	ACUSON Sequoia Ultrasound System	Siemens Healthineers	Rely+On PeraSafe	Tristel Trio Wipe, Tristel Duo ULT	
BK Medical	bkSpecto	KeboMed AS	Rely+On PeraSafe	Tristel Trio Wipe, Tristel Duo ULT	Desinfeksjonssprit 70 - 75 % <i>kun til ikke kritisk utstyr</i>
FUJIFILM SonoSite	SonoSite	Secma AS		Tristel Trio Wipe, Tristel Duo ULT	
GE Healthcare, GE Vingmed Ultrasound	GE Healthcare	GE Healthcare Norge AS, GE Vingmed Ultrasound	Rely+On PeraSafe	Tristel Trio Wipe, Tristel Duo ULT	Desinfeksjonssprit 70 - 75 % <i>kun til ikke kritisk utstyr</i>
Philips	Philips ATL	Philips Norge AS		Tristel Trio Wipe, Tristel Duo ULT	
Hitachi – Aloka	Aloka	Scan-Med AS Norway	Rely+On PeraSafe	Tristel Trio Wipe, Tristel Duo ULT	
Canon Medical	Canon Aplio, Xario, Aviamo	Tromp Medical AS	Rely+On PeraSafe, Clinell Peracetic Acid Wipes	Tristel Trio Wipe, Tristel Duo ULT	Desinfeksjonssprit 70 - 75 % <i>kun til ikke kritisk utstyr</i>
Samsung Healthcare	Samsung	Inter Medical AS		Tristel Trio Wipe, Tristel Duo ULT LifeClean	
Natus Neuro	Natus® InVisus™ Ultrasound Sonara	Timik AS		Tristel Trio Wipe, Tristel Duo ULT	